CE59PT0402

Hangzhou Clongene Biotech Co., Ltd.

Risk Analysis Report

COVID-19 Antigen Rapid Test Cassette (Nasal Swab)

| Revision | Ву | Date | |
|----------|---------------|------------|--|
| 1.0 | Xingqiao Pang | 29/01/2021 | |
| | | | |
| | | | |
| | | | |

| Hangzhou Clongene Biotech Co., Ltd. | Number: CE59PT0402 | |
|---------------------------------------------------|--------------------|--|
| COVID-19 Antigen Rapid Test Cassette (Nasal Swab) | Page:1/16 | |
| Risk Analysis Report | Rev. 1.0 | |

1. Summarize

This report is a risk analysis to the COVID-19 Antigen Rapid Test Cassette (Nasal Swab). The report would determine all potential hazards and the potential reason that each hazard would cause, and estimate the probability of hazard occurrences, and, simultaneously, estimate the damage level once hazards occur. If there are the measures to reduce the risk, expatiate it, and estimate the residual risk levels.

Conclusion: By taking the adaptable measures, all risk that would cause hazard would be reduced to the level that can be as low as reasonably practicable, that is to say, the potential risk would be reduced to the practicable levels.

2. Purpose

The purpose of risk analysis is to estimate the hazard, which is took place in the use of the COVID-19 Antigen Rapid Test Cassette (Nasal Swab).

3. Reference standard

- a) EN ISO 14971:2019 Medical devices-Application of risk management to medical devices
- b) EN 13641:2002 Elimination or reduction of risk of infection related to in vitro diagnostic reagents
- c) 98/79/EC Directive 98/79/EC of the European Parliament and of the council of 27 October 1998 on in vitro diagnostic medical devices
- d) EN ISO 13485:2016 Medical devices Quality management systems Requirements for regulatory purposes
- e) ISO 17511:2020 In vitro diagnostic medical devices—Requirements for establishing metrological traceability of values assigned to calibrators, trueness control materials and human samples
- f) EN ISO 18113-1:2011 In vitro diagnostic medical devices Information supplied by the manufacturer (labelling) Part 1: Terms, definitions and general requirements
- g) EN ISO 18113-2:2011 In vitro diagnostic medical devices Information supplied by the manufacturer (labelling) - Part 2: In vitro diagnostic reagents for professional use
- h) EN ISO 15223-1:2016 Medical devices Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements
- i) EN 13975:2003 Sampling procedures used for acceptance testing of in vitro diagnostic medical devices — Statistical aspects
- j) EN 13612: 2002 Performance evaluation of in vitro diagnostic medical devices
- k) EN ISO 23640:2015 In vitro diagnostic medical devices Evaluation of stability of in vitro diagnostic reagents (ISO 23640:2011)
- 1) EN 62366-1: 2015 Medical devices Part1: Application of usability engineering to medical devices

4. Production specification: Structure and principal

The COVID-19 Antigen Rapid Test Cassette (Nasal Swab) is designed to detect nucleocapsid antigen

| Hangzhou Clongene Biotech Co., Ltd. | Number: CE59PT0402 |
|---------------------------------------------------|--------------------|
| COVID-19 Antigen Rapid Test Cassette (Nasal Swab) | Page:2/16 |
| Risk Analysis Report | Rev. 1.0 |

from the SARS-CoV-2 in nasal swab from patients who are suspected of COVID-19 by their healthcare provider.

Principle: The COVID-19 Antigen Rapid Test Cassette (Nasal Swab) is a lateral flow immunoassay based on the principle of the double-antibody sandwich technique. SARS-CoV-2 nucleocapsid protein monoclonal antibody conjugated with color micro- particles is used as detector and sprayed on conjugation pad.

During the test, SARS-CoV-2 antigen in the specimen interacts with SARS-CoV-2 antibody conjugated with color microparticles making antigen-antibody labeled complex. This complex migrates on the membrane via capillary action until the test line, where it will be captured by the pre-coated SARS- CoV-2 nucleocapsid protein monoclonal antibody. A colored test line (T) would be visible in the result window if SARS-CoV-2 antigens are present in the specimen. Absence of the T line suggests a negative result. The control line (C) is used for procedural control, and should always appear if the test procedure is performed properly.

5. Risk analysis group

Risk analysis group is composed of members as follows:

| Title | Name | Project |
|----------------------------------|---------------|---------------------------------------------------|
| Project Team Leader | Xingqiao Pang | Whole plan of risk analysis |
| R & D Department Manager | Qingqing Chen | Evaluate the harm degree as a view of biological |
| Quality Department Manager | Xingqiao Pang | Evaluate the harm degree as a view of doctor |
| Manufacturing Department Manager | Jun Wu | Evaluate the harm degree as a view of chemist and |
| | | mechanical performance |
| Marketing Department Manager | Guoqin Chen | Evaluate the harm degree as a view of users |

6. Identification of qualitative and quantitative characteristics (acc. to EN ISO 14971:2019)

a) What is the intended use and how to be used?

The COVID-19 Antigen Rapid Test Cassette (Nasal Swab) is a lateral flow immunoassay intended for the qualitative detection SARS-CoV-2 nucleocapsid antigens in nasal swab from individuals who are suspected of COVID-19 by their healthcare provider.

The COVID-19Antigen Rapid Test Cassette (Nasal Swab) is intended for use by medical professionals or trained operators who are proficient in performing lateral flow tests. The product may be used in any laboratory and non-laboratory environment that meets the requirements specified in the Instructions for Use and local regulation.

The method of using:

Note: Allow the test cassettes, reagents and specimens to equilibrate to room temperature $(15-30^{\circ}C \text{ or } 59-86^{\circ}F)$ prior to testing.

| Hangzhou Clongene Biotech Co., Ltd. | Number: CE59PT0402 |
|---------------------------------------------------|--------------------|
| COVID-19 Antigen Rapid Test Cassette (Nasal Swab) | Page:3/16 |
| Risk Analysis Report | Rev. 1.0 |

- 1. Put an extraction tube on the work station.
- 2. Unscrew the lid of an extraction reagent. Add all of the extraction reagent into the extraction tube.
- 3. Sampling refers to section 'Specimen Collection'.
- 4. Insert the nasal swab specimen into the extraction tube which contains extraction reagent. Roll the swab at least 5 times while pressing the head against the bottom and side of the extraction tube. Leave the nasal swab in the extraction tube for one minute.
- 5. Remove the nasal swab while squeezing the sides of the tube to extract the liquid from the swab. The extracted solution will be used as test sample.
- 6. Cover the extraction tube with a dropper tip tightly.



- 7. Remove the test cassette from the sealed pouch.
- 8. Reverse the specimen extraction tube, holding the tube upright, transfer 3 drops (approximately 100μ L) slowly to the specimen well (S) of the test cassette, then start the timer.
- 9. Wait for colored lines to appear. Interpret the test results at 15 minutes. Do not read results after 20 minutes.



b) Intended to be in contact with the patient or other persons?

The COVID-19 Antigen Rapid Test Cassette (Nasal Swab) which is intended for in vitro diagnostic use only will not contact the patient or other persons.

c) Materials/components used

All materials or components used in the reagents had been widely used in many IVD products, which have been used in clinical test for several years and proven on the safe side. And many of those IVD products have the CE certifications.

d) Is energy delivered to or extracted from the patient?

There is not any energy flowing in or out from the patient.

| Hangzhou Clongene Biotech Co., Ltd. | Number: CE59PT0402 | |
|---------------------------------------------------|--------------------|--|
| COVID-19 Antigen Rapid Test Cassette (Nasal Swab) | Page:4/16 | |
| Risk Analysis Report | Rev. 1.0 | |

e) Are substances delivered to or extracted from the patient?

There are not any substances coming into the patients bodies. The nasal swab is the substance out from the patients. In general, the specimen will be collected by trained clinical laboratory personnel specifically instructed and trained in vitro diagnostic procedures following standard laboratory procedures, which would not make any hazard for patients.

f) Are biological materials processed by the reagent for subsequent re-use, transfusion or transplantation?

To prevent inaccurate test results due to improper specimen handling and/or transportation, after collecting specimens in accordance with routine clinical laboratory procedures, we need to store the swab in the extraction reagent provided with the kit, or immerse the swab head in the prescribed virus preservation solution to store. In addition, we need to extract the solution from the collected swab specimen as the test sample through the extraction tube with the extraction reagent added. Therefore, it is strongly recommended to train in specimen collection.

- g) Is the reagent supplied sterile or intended to be sterilized by the user?The reagents do not need the asepsis style or sterilization.
- h) Is the reagent intended to modify the patient environment?The reagents are not intended to modify the patient environment.
- i) If the reagent has the measure function or analysis function? The reagents have not measure function or analysis function.
- j) Is the reagent intended for use in conjunction with other medical devices, medicines or other medical technologies?

The reagents do not use in conjunction with any medicines or other medical technologies.

- k) Are there unwanted outputs of energy or substances?The reagents have not any energy or matter output unexpected.
- 1) Is the reagent susceptible to environmental influences?

The condition for transporting and storing the reagents, which are not special requests, as follows: the temperature range is $4^{\circ}C \sim 30^{\circ}C$, the relative humidity is normal.

m) Does the reagent influence the environment?

All nasal swabs specimens and used reagents should be considered potentially infectious and avoided contact with skin.

n) Are there essential consumables or accessories associated with the reagent?

| Hangzhou Clongene Biotech Co., Ltd. | Number: CE59PT0402 | |
|---------------------------------------------------|--------------------|--|
| COVID-19 Antigen Rapid Test Cassette (Nasal Swab) | Page:5/16 | |
| Risk Analysis Report | Rev. 1.0 | |

The accessories going with the reagents are dropper tip, extraction tube, work station and sterilized swab.

o) Is maintenance or calibration necessary?

Each reagent is used only once. So the reagents need not maintenance and/or calibration.

- p) Does the reagent contain software?The reagent does not contain software.
- q) Does the reagent have a restricted shelf-life?The reagents have the shelf-life of 24 months.
- r) Are there any delayed or long-term use effects?
 Test results of the reagent may be unauthentic if lingering and/or long-term used.
- s) What determines the lifetime of the reagent?

The lifetime of the reagent is influenced by the environment temperature and the seal of the foil pouch. When the environment temperature exceeds 30° C, or the foil pouch was seal incompletely, the lifetime of the reagent will be reduced greatly.

- t) Is the reagent intended for single use? Yes.
- u) Is safe decommissioning or disposal of the medical reagent necessary?
 Discard the reagent in the bio-hazardous dustbin after single use.
- v) Does installation or use of the reagent require special training or special skills? The operator must be pre-trained specially before use the reagent.
- w) How will information for safe use be provided?

The information for safe use included in the labels and IFU will be provided directly to the end user.

- x) Will new manufacturing processes need to be established or introduced? Any new manufacturing processes needn't be established or introduced.
- y) Is successful application of the medical device critically dependent on human factors? The user should read "Test Procedure" in the IFU before performing the test.

7. Risk assessment of hazards

Severity

| Ranking | Ranking (descriptive) | Description |
|---------|-----------------------|-------------|
|---------|-----------------------|-------------|

| Hangzhou Clongene Biotech Co., Ltd. | Number: CE59PT0402 | |
|---------------------------------------------------|--------------------|--|
| COVID-19 Antigen Rapid Test Cassette (Nasal Swab) | Page:6/16 | |
| Risk Analysis Report | Rev. 1.0 | |

| 1 | Negligible | Temporary discomfort or inconvenience to user/operator |
|---|--------------|------------------------------------------------------------------|
| 2 | Marginal | Injury (usually requiring medical intervention) to user/operator |
| 3 | Critical | Severe injury or impairment to user/operator |
| 4 | Catastrophic | Life-threatening or death to user/operator |

Probability of Occurrence

| Ranking | Likelihood | Description | Probability |
|---------|------------|--------------------------------|------------------------|
| 1 | Improbable | Unlikely to occur | <10-6 |
| 2 | Remote | Unlikely but possible to occur | $10^{-4} \sim 10^{-6}$ |
| 3 | Occasional | Likely to occur sometime | $10^{-3} \sim 10^{-4}$ |
| 4 | Probable | Likely to occur several times | $10^{-1} \sim 10^{-3}$ |
| 5 | Frequent | Likely to occur frequently | ≥10 ⁻¹ |

Risk Assessment Table

| | | Severity | | | |
|-------------|------------|------------|----------|----------|--------------|
| | | Negligible | Marginal | Critical | Catastrophic |
| | Frequent | | | | |
| | Probable | | | | |
| Probability | Occasional | | | | |
| | Remote | | | | |
| | Improbable | | | | |

Risk Assessment Levels

| Level | Safety Risk | Description |
|------------------|-------------|-------------------------------------------------------------------|
| Red Unacceptable | | Risk in this region is not tolerated. |
| | | Resolution Required – redesign, do not release. |
| Green | Accortable | Risk is considered to be negligible compared to the risk of other |
| | Acceptable | hazards. |

| Hangzhou Clongene Biotech Co., Ltd. | Number: CE59PT0402 | | | |
|---------------------------------------------------|--------------------|--|--|--|
| COVID-19 Antigen Rapid Test Cassette (Nasal Swab) | Page:7/16 | | | |
| Risk Analysis Report | Rev. 1.0 | | | |

| | Risk Assessment before Mitigation | | | | | | Risk Ass | essment after | Mitigation |
|---------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------|----------|------------------------------|----------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------|------------------------------|------------|
| Hotord | | Risk Estimation | | Dick | Bick Control Moscuro | Risk E | stimation | Bick | |
| Hazaro | Cause | Adverse Effect | Severity | Probability of Occurrence | Evaluation | Risk Control Measure | Severity | Probability of Occurrence | Evaluation |
| False Results | The quality of the antibodies does not produce the expected performance. | Users requires medical attention | 3 | 3 | Unacceptable | Design control phase 3 & 4 reports require verification and validation. Incoming QC inspection documents -Specification For 101005001 (QC-2050) -Specification For 101005301 (QC-2053) -Specification For 101000401 (QC-2004) | 3 | 1 | Acceptable |
| False Results | Critical raw material (mouse- anti-goat IgG polyclonal antibody, SARS-CoV-2 nuclenocapsid protein monoclonal antibody, SARS-CoV-3 nuclenocapsid protein monoclonal antibody) is non-conforming. | Users requires medical attention | 3 | 3 | Unacceptable | Design control phase 3 & 4 reports require product testing, verification & validation; incoming QC inspection must be done. -Specification For 101005001 (QC-2050) -Specification For 101005301 (QC-2053) -Specification For 101000401 (QC-2004) | 3 | 1 | Acceptable |
| False Results | Inadequate stability of the raw material | Users requires medical attention | 3 | 3 | Unacceptable | Qualification of alternate suppliers if possible, appropriate QC inspection criteria. Stability testing should be done. -Specification of QC -Procurement Specification | 3 | 1 | Acceptable |
| False Results | The sample pad does not conform to specifications. | Users requires medical attention | 3 | 3 | Unacceptable | Sample pad inspection procedure -Inspection Criterion For Fiberglass (QC-3003) | 3 | 1 | Acceptable |

| Hangzhou Clongene Biotech Co., Ltd. | Number: CE59PT0402 | | |
|---------------------------------------------------|--------------------|--|--|
| COVID-19 Antigen Rapid Test Cassette (Nasal Swab) | Page:8/16 | | |
| Risk Analysis Report | Rev. 1.0 | | |

| | | Risk Assessment be | | Risk Ass | essment after | Mitigation | | | |
|---------------|---------------------------------------------------------------------|----------------------------------|----------|------------------------------|---------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------|------------------------------|------------|
| Hotord | | | Risk E | stimation | Bick | Bick Control Moscuro | Risk E | stimation | Bick |
| ΠαΖαΓΟ | Cause | Adverse Effect | Severity | Probability of Occurrence | Evaluation | Risk Control Measure | Severity | Probability of Occurrence | Evaluation |
| False Results | Packaging materials do not conform to specifications. | Users requires medical attention | 3 | 3 | Unacceptable | Packaging materials Inspection and testing control procedure -Inspection Criterion For Carton (QC-3023) -Inspection Criterion For Foil Bag (QC-3026) | 3 | 1 | Acceptable |
| False Results | The storage conditions of raw material are not correct. | Users requires medical attention | 3 | 3 | Unacceptable | The shelf life of the raw material should be defined in COA from vender and validated by R&D during the product development; Manufacturing SOPs -Raw Material Warehouse Management System (QMS-P01) | 3 | 1 | Acceptable |
| False Results | Operator not following manufacturing procedures. | Users requires medical attention | 3 | 3 | Unacceptable | manufacturing & QC inspection documents; training procedures -Specification of QC -Procurement Specification -Standard Operating Procedures -Control Procedure of Manufacturing Process(QMS-L00) -Training Management System (QMS-D02) | 3 | 1 | Acceptable |
| False Results | Lack of reagent integrality (e.g. the desiccant is missing) | Users requires medical attention | 3 | 3 | Unacceptable | QC inspection documents -Standard Practice for assembly of drop reagent (MF-2021) - Functional Testing Procedure | 1 | 1 | Acceptable |

| Hangzhou Clongene Biotech Co., Ltd. | Number: CE59PT0402 |
|---------------------------------------------------|--------------------|
| COVID-19 Antigen Rapid Test Cassette (Nasal Swab) | Page:9/16 |
| Risk Analysis Report | Rev. 1.0 |

| | Risk Assessment before Mitigation | | | | | | Risk Ass | essment after | Mitigation |
|----------------|----------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------|----------|------------------------------|--------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------|------------------------------|------------|
| Hazard | | | Risk E | stimation | Rick | Risk Control Measure | Risk E | stimation | Rick |
| ΠαΖάιυ | Cause | Adverse Effect | Severity | Probability of Occurrence | Evaluation | Kisk Control Measure | Severity | Probability of Occurrence | Evaluation |
| False Results | Improper package | Users requires medical attention | 3 | 3 | Unacceptable | For COVID-19 Semi-Finished Product - Preparation Of COVID-19 Control Standards - Final Testing Procedure For COVID-19 Product Packaging SOP and finished product QC - Final Testing Procedure For | 3 | 1 | Acceptable |
| False Results | Incorrect QC Samples used, inappropriate QC procedure | Users requires medical attention | 3 | 3 | Unacceptable | COVID-19 Product Manufacturing & QC inspection documents; training procedures QC inspection documents; training procedures -Specification of QC -Procurement Specification - Functional Testing Procedure For COVID-19 Semi-Finished Product - Final Testing Procedure For COVID-19 Product -Training Management System (QMS-D02) | 3 | 1 | Acceptable |
| False-positive | The nature/ characteristic of the sample gives interference | Can have extremely devastating psychological effects on the person tested ,or damage to the reputation, even lead to unemployment and | 3 | 3 | Unacceptable | Interferences substances were tested during development and field trials The IFU indicates which samples to use. -Instruction for use of COVID-19 Antigen Rapid | 3 | 1 | Acceptable |

| Hangzhou Clongene Biotech Co., Ltd. | Number: CE59PT0402 | | | |
|---------------------------------------------------|--------------------|--|--|--|
| COVID-19 Antigen Rapid Test Cassette (Nasal Swab) | Page:10/16 | | | |
| Risk Analysis Report | Rev. 1.0 | | | |

| | Risk Assessment before Mitigation | | | | | | Risk Assessment after Mitigation | | |
|----------------|-----------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------|------------------------------|--------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------|------------------------------|------------|
| Hazard | | | Risk E | stimation | Dick | Bisk Control Mossuro | Risk E | stimation | Pick |
| падаги | Cause | Adverse Effect | Severity | Probability of Occurrence | Evaluation | Risk Control Measure | Severity | Probability of Occurrence | Evaluation |
| | | serious medical consequences | | | | Test Cassette (Nasal Swab) (CE59-attachment 7.2) Validation studies - Interference Study Report (CE59PT0507) | | | |
| False-positive | Cross reaction with other interference substances | Can have extremely devastating psychological effects on the person tested ,or damage to the reputation, even lead to unemployment and serious medical consequences | 3 | 3 | Unacceptable | Interferences substances were tested during development and field trials The IFU indicates which samples to use. - Instruction for use of COVID-19 Antigen Rapid Test Cassette (Nasal Swab) (CE59-attachment 7.2) Validation studies - Analytical Specificity (Cross-reactivity) Study Report (CE59PT0503) | 3 | 1 | Acceptable |
| False-negative | Interferences due to the different types of samples | Physical harm caused to the person tested, and lead to some adverse effect on society | 3 | 3 | Unacceptable | The different samples to use are tested during development and field trials. The IFU indicates which samples to use. - Instruction for use of COVID-19 Antigen Rapid Test Cassette (Nasal Swab) (CE59-attachment 7.2) Validation studies | 3 | 1 | Acceptable |
| False Results | Storage and transport conditions are not correct | Users requires medical attention | 3 | 3 | Unacceptable | Labels and IFU provide the proper storage conditions. -Labels and Instruction for use of COVID-19 Antigen | 3 | 1 | Acceptable |

| Hangzhou Clongene Biotech Co., Ltd. | Number: CE59PT0402 | | | |
|---------------------------------------------------|--------------------|--|--|--|
| COVID-19 Antigen Rapid Test Cassette (Nasal Swab) | Page:11/16 | | | |
| Risk Analysis Report | Rev. 1.0 | | | |

| | | Risk Assessment be | | Risk Ass | essment after | Mitigation | | | |
|----------------|-------------------------|-------------------------|----------|------------------------------|---------------|----------------------------------------------------|----------|------------------------------|--------------|
| Hotord | | | Risk E | stimation | Dick | Bick Control Moscuro | Risk E | stimation | Bick |
| пазаго | Cause | Adverse Effect | Severity | Probability of Occurrence | Evaluation | Risk Control Measure | Severity | Probability of Occurrence | Evaluation |
| | | | | | | Rapid Test Cassette (Nasal | | | |
| | | | | | | Swab) (CE59-attachment 7.1, | | | |
| | | | | | | 7.2) | | | |
| | | | | | | Validation studies | | | |
| | | | | | | - Stability Study Report (CE59PT0501) | | | |
| | | | | | | Labels and IFU provide the | | | |
| | | | | | | proper intended use. | | | |
| Falsa Pasulta | Deviate from the | Users requires medical | 3 | 2 | Unacceptable | -Labels and Instruction for use | 2 | 1 | Accortable |
| Faise Results | intended use | attention | 5 | 3 | | of COVID-19 Antigen Rapid | 5 | 1 | Acceptable |
| | | | | | | Test Cassette (Nasal Swab) | | | |
| | | | | | | (CE59-attachment 7.1, 7.2) | | | |
| | Finger contact leads to | | 3 | | | The operator is a trained | | | |
| | | | | 3 | | professional. The IFU warns | | | |
| | | | | | | the operator not to touch the | | | |
| False Results | contamination of the | Users requires medical | | | Unacceptable | membrane. | 3 | 1 | Acceptable |
| | test | attention | | | | - Instruction for use of COVID 10 Antigan Panid | | | |
| | | | | | | Test Cassette (Nasal Swah) | | | |
| | | | | | | (CE59-attachment 7.2) | | | |
| | | | | | | Design control phase 3 & 4 | | | |
| | | | | | | reports require verification and | | | |
| | | Physical harm caused to | | | | validation and Information in | | | |
| | Insufficient sample | the person tested, and | 2 | 2 | Uncocontoble | IFU | 2 | 1 | A accentable |
| Faise-negative | applied | lead to some adverse | 3 | 5 | Unacceptable | - Instruction for use of | 3 | 1 | Acceptable |
| | | effect on society | | | | COVID-19 Antigen Rapid | | | |
| | | | | | | Test Cassette (Nasal Swab) | | | |
| | | | | | | (CE59-attachment 7.2) | | | |
| | | Users requires medical | | | | The IFU describes the correct | | | |
| False Results | Improper operation | attention | 3 | 3 | Unacceptable | test procedure and stresses the | 3 | 1 | Acceptable |
| | | | | | | critical procedure. | | | |

| Hangzhou Clongene Biotech Co., Ltd. | Number: CE59PT0402 | | | |
|---------------------------------------------------|--------------------|--|--|--|
| COVID-19 Antigen Rapid Test Cassette (Nasal Swab) | Page:12/16 | | | |
| Risk Analysis Report | Rev. 1.0 | | | |

| | | Risk Assessment be | | Risk Assessment after Mitigation | | | | | |
|----------------------|-----------------------------------------------------------------------------------------|----------------------------------|----------|----------------------------------|--------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------|------------------------------|------------|
| Hazard | | | Risk E | stimation | Dick | Bick Control Moscuro | Risk Estimation | | Pick |
| Hazaru | Cause | Adverse Effect | Severity | Probability of Occurrence | Evaluation | Risk Control Measure | Severity | Probability of Occurrence | Evaluation |
| | | | | | | - Instruction for use of COVID-19 Antigen Rapid Test Cassette (Nasal Swab) (CE59-attachment 7.2) | | | |
| Incorrect Reading | Improper operating environment (e.g. Ambient temperature too high or too low) | Users requires medical attention | 3 | 3 | Unacceptable | Appropriate operating environment are established during development and validation. And the IFU mentions the conditions. - Instruction for use of COVID-19 Antigen Rapid Test Cassette (Nasal Swab) (CE59-attachment 7.2) | 3 | 1 | Acceptable |
| Incorrect Reading | Instructions for use regarding Read Result Time not followed. | Users requires medical attention | 3 | 3 | Unacceptable | The read result times are established during development and validated during the validation studies. The IFU describes the correct reading time. - Instruction for use of COVID-19 Antigen Rapid Test Cassette (Nasal Swab) (CE59-attachment 7.2) | 3 | 1 | Acceptable |
| Incorrect Reading | Confusion between control line and test line results | Users requires medical attention | 3 | 3 | Unacceptable | Information presented in the IFU - Instruction for use of COVID-19 Antigen Rapid Test Cassette (Nasal Swab) (CE59-attachment 7.2) | 3 | 1 | Acceptable |
| Users contamination | Biological and/or chemical contamination | Health Risk | 3 | 3 | Unacceptable | The operator is a trained professional. The package insert warns the operator about | 3 | 1 | Acceptable |

| Hangzhou Clongene Biotech Co., Ltd. | Number: CE59PT0402 | | |
|---------------------------------------------------|--------------------|--|--|
| COVID-19 Antigen Rapid Test Cassette (Nasal Swab) | Page:13/16 | | |
| Risk Analysis Report | Rev. 1.0 | | |

| | | Risk Assessment b | | Risk Assessment after Mitigation | | | | | |
|-----------------------------|----------------------------------------------------------------------------------------|------------------------------------|----------|----------------------------------|--------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------|------------------------------|------------|
| Horord | | | Risk E | stimation | Pick | Bick Control Moscuro | Risk Estimation | | Bick |
| Hazaru | Cause | Adverse Effect | Severity | Probability of Occurrence | Evaluation | Risk Control Measure | Severity | Probability of Occurrence | Evaluation |
| | | | | | | any potentially hazardous material and the necessity to wear gloves. - Instruction for use of COVID-19 Antigen Rapid Test Cassette (Nasal Swab) (CE59-attachment 7.2) | | | |
| Users contamination | The operator is in contact with the sample during or after sample application | Health Risk | 3 | 3 | Unacceptable | The operator is a trained professional. The package insert warns the operator that the sample is potentially infectious. The package insert mentions the use of gloves -Labels and Instruction for use of COVID-19 Antigen Rapid Test Cassette (Nasal Swab) (CE59-attachment 7.1, 7.2) | 3 | 1 | Acceptable |
| False Results | Unqualified samples are used | Users need more medical testing | 3 | 4 | Unacceptable | The requirements of sample collection and the storage conditions for sample are established during development and validation. This information will be mentioned in package insert. - Instruction for use of COVID-19 Antigen Rapid Test Cassette (Nasal Swab) (CE59-attachment 7.2) | 3 | 1 | Acceptable |
| Environmental contamination | Improper treatment the waste/appliance | Health Risk | 3 | 4 | Unacceptable | The IFU warns to handle all specimens and other castoff as potentially infectious. | 3 | 1 | Acceptable |

| Hangzhou Clongene Biotech Co., Ltd. | Number: CE59PT0402 |
|---------------------------------------------------|--------------------|
| COVID-19 Antigen Rapid Test Cassette (Nasal Swab) | Page:14/16 |
| Risk Analysis Report | Rev. 1.0 |

| Hazard | Risk Assessment before Mitigation | | | | | Risk Assessme | | essment after | Mitigation |
|--------|-----------------------------------|----------------|-------------------------|-------------------------|------|----------------------------|-----------------|---------------|------------|
| | | Adverse Effect | Risk Estimation | | Risk | Risk Control Measure | Risk Estimation | | Pick |
| | Cause | | Soverity Probability of | Soverity Probability of | | | Evaluation | | |
| | | | Seventy | Occurrence | | | Seventy | Occurrence | |
| | | | | | | - Instruction for use of | | | |
| | | | | | | COVID-19 Antigen Rapid | | | |
| | | | | | | Test Cassette (Nasal Swab) | | | |
| | | | | | | (CE59-attachment 7.2) | | | |

| Hangzhou Clongene Bi | otech Co., Ltd. | Number: CE59PT0402 | |
|-----------------------------|-------------------------|--------------------|-----|
| COVID-19 Antigen Rapid Test | t Cassette (Nasal Swab) | Page:15/16 | |
| Risk A | | nalysis Report | Rev |

8. Conclusion

When we use some measurement to solve or reduce all the potential risks of the reagent, effect is obvious, as showing above. It can be seen that the residual risks are as low as reasonably practicable and is located in the area of comprehensive acceptable, so risk and advantage is proportional. The COVID-19 Antigen Rapid Test Cassette (Nasal Swab) has used for long time for all over the world, and that advantage is greater than risk is obvious.

As it is has been seen that the residual risk of the COVID-19 Antigen Rapid Test Cassette (Nasal Swab) is reasonably practicable, and the COVID-19 Antigen Rapid Test Cassette (Nasal Swab) is safe and effective in clinic.

9. Reference

MAUDE Database: http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfMAUDE/search.CFM

| Hangzhou Clongene Biotech Co., Ltd. | Number: CE59PT0402 |
|-------------------------------------|--------------------|
| Dick And Link D | Page:16/16 |
| Risk Analysis Report | Rev. 1.0 |

The Risk Control Measure Effectiveness and Verification Methods Effectiveness have been reviewed and all hazards / risks are either acceptable or ALARP

| Product Identification: | COVID-19 Antigen Rapid Test Cassette (Nasal Swab) |
|--------------------------------|---------------------------------------------------|
|--------------------------------|---------------------------------------------------|

Team Leader:

-

Xingqiao Pang

Risk Analysis Team Membership

Date: 2021-01-29

| Attendees: | Printed Name | Signature | Department |
|------------|---------------|-----------|---------------|
| | Lulu Zhang | 3 Ports | R&D |
| | Qingqing chen | 降、素素 | R&D |
| | Jun Wu | 4103 | Manufacturing |
| | Xingqiao Pang | 45 5 1 | Quality |
| | Guoqin Chen | PORA | Marketing |
| | |) | |

Design Review Committee Disposition

| Department | Recommend Approval | Printed Name | Signature | Date |
|---------------|-----------------------------------|---------------|-----------|------------|
| R&D | [√] Yes [] No [] Cond. | Qingqing chen | 游客东 | 171.121 |
| Manufacturing | $[\checkmark]$ Yes [] No [] Cond. | Jun Wu | 4/4 | 20101.01 |
| Quality | [J Yes [] No [] Cond. | Xingqiao Pang | Its I ~ | 10, 10, 10 |
| Marketing | [🖌 Yes [] No [] Cond. | Guoqin Chen | POBS | 24.51.4 |
| | | | | 2021.212 |

Risk Analysis Report Final Disposition (The final disposition is ruled by majority vote of the Design Review Committee. This portion is filled out by the Regulatory representative.)

Approved

- □ Approved, but requires follow-up (see comments)
- □ Not approved, requires additional work (see comments)

Comments:_____

Approved By:

Ť

Date: Jon S. Y